

# NuMega Resonance Labs

*Fast Reliable Services*

11526 Sorrento Valley Road, Suite B-2 ♦ San Diego, CA 92121 ♦ USA  
Phone 858-793-6057 ♦ Fax 858-793-2607

October 24, 2023

Jiong Zou  
Dexcom, Inc  
6350 Sequence Drive  
San Diego, CA 92121  
[jzou@dexcom.com](mailto:jzou@dexcom.com)

NuMega Reference #

2309BROM.245D

## Quantitative Analysis Report

*Bruker 500 MHz, AVII-500 and AV-500*

Sample ID	ABT-089 HCl
Standard	Dimethylsulfone
Weight of Standard (wt <sub>st</sub> )	0.992
Weight of Sample (wt <sub>sa</sub> )	8.847
Sample Molecular Weight (MW <sub>sa</sub> )	265.18
Standard Molecular Weight (MW <sub>st</sub> )	94.13
Integration of Sample (I <sub>sa</sub> )	104.82
Integration of Standard (I <sub>st</sub> )	100
# of Equivalent Nuclei in Standard (n <sub>st</sub> )	6
# of Equivalent Nuclei in Sample (n <sub>sa</sub> )	2
Assigned Purity of Standard % (P)	99.82
Purity %	99.2%

Formula Used for Calculation: 
$$WT\% = \frac{wt_{st}}{wt_{sa}} \times \frac{MW_{sa}}{MW_{st}} \times \frac{I_{sa}}{I_{st}} \times \frac{n_{st}}{n_{sa}} \times P$$