

NuMega Resonance Labs

Fast Reliable Services

11526 Sorrento Valley Road, Suite B-2 ♦ San Diego, CA 92121 ♦ USA
Phone 858-793-6057 ♦ Fax 858-793-2607

August 13, 2026

NuMega Reference #
2604BROM.932F

Company: Bromantane LLC

Quantitative Analysis Report

Bruker 500 MHz, AVII-500 and AV-500

Sample ID	Sialic Acid
Standard	Ethylene Carbonate
Weight of Standard (wt _{st})	1.931
Weight of Sample (wt _{sa})	9.361
Assumed Sample Molecular Weight (MW _{sa})	309.27
Standard Molecular Weight (MW _{st})	88.06
Integration of Sample (I _{sa})	32.30
Integration of Standard (I _{st})	100
# of Equivalent Nuclei in Standard (n _{st})	4
# of Equivalent Nuclei in Sample (n _{sa})	1
Assigned Purity of Standard % (P)	99.71
Purity %	93.3

Formula Used for Calculation:

$$\text{Purity \%} = \frac{wt_{st}}{wt_{sa}} \times \frac{MW_{sa}}{MW_{st}} \times \frac{I_{sa}}{I_{st}} \times \frac{n_{st}}{n_{sa}} \times P$$